



CERTIFICATE OF TEST

COMPANY: IFD LLC / DBA Desin LLC
714 E. Monument Avenue
Dayton, OH 44402 USA

PRODUCT NAME: Robotic Dining / Eating Device

MODEL NO(S): Obi

TEST RESULTS: In compliance

REPORT NO(S): F2P22725A-01i; F2P22725A-01S; F2P22725A-04S

STANDARD(S): IEC/EN 60529: 2013, Degrees of protection provided by enclosures (IP21)
IEC 60601-1:2005, COR1:2006, COR2:2007, AMD1:2012 (or IEC 60601-1:2012 reprint) Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential Performance
IEC 60601-1-11:2015 (Second Edition) for use in conjunction with IEC 601:2005 (Third Edition) + A1:2012 Medical Electrical Equipment – Part 1-11: General Requirements for Basic Safety and Essential Performance – Collateral Standard: Requirements for Medical Electrical Equipment and Medical Electrical Systems used in the Home Healthcare Environment

TEST LABORATORY: F2 Labs
26501 Ridge Road,
Damascus, MD 20872 USA

Signature:

Name:

Andrew Bagnall

Title:

Vice President

Date: 2020-03-18





CERTIFICATE OF TEST



On-Site Compliance
Testing